

Cytometer Performance Report

Cytometer: FACSCantoII
Cytometer Name: FACSCantoII
Serial Number: V33896201821
Input Device: Carousel
Tube Loaded Manually: Yes
Cytometer Configuration: Euroflow 3-laser, 8-color (4-2-2)

User: Administrator
Institution: N/A
Software: BD FACSDiva 8.0.2
Date: 10/03/2018 11:35 AM
Cytometer Baseline: 09/25/2018 11:17 AM
P/F: Pass

Setup Beads

Bead Product: CST Setup Beads **Part #:** 910859
Lot ID: 80291 **Expiration Date:** 02/29/2020
Bead Lot Information: Available

Detector Settings

Laser	Detector	Parameter	Target Value	Actual Target Value	% Difference Target Value	Bright Bead % Robust CV	Mid Bead Median Channel	Mid Bead % Robust CV
Blue	FSC	FSC	125000	123520	-2	0.9	123590	0.89
Blue	F	SSC	125000	125974	0	1.69	126748	1.72
Blue	E	FITC	16596	16234	-3	2.79	349	21.32
Blue	D	PE	29392	28744	-3	1.84	624	16.04
Blue	B	PerCP-Cy5-5	38531	38481	-1	3.98	1176	21.59
Blue	A	PE-Cy7	26351	26497	0	7.16	651	42.02
Red	C	APC	43035	42296	-2	2.67	1594	13.26
Red	A	APC-Cy7	25611	25263	-2	2.8	871	12.95
Violet	B	V450	8803	8490	-4	4.2	467	14.45
Violet	A	V500	32560	32050	-2	3.9	748	10.96

Detector Settings (Continued)

Laser	Detector	Parameter	Dim Bead Median Channel	Dim Bead % Robust CV	PMTV	Δ PMTV	Qr	Br	P/F
Blue	FSC	FSC	20400	2.56	602	5	N/A	N/A	Pass
Blue	F	SSC	57205	1.12	568	13	N/A	N/A	Pass
Blue	E	FITC	61	68.71	554	4	0.0171	139	Pass
Blue	D	PE	130	52.83	553	4	0.1284	230	Pass
Blue	B	PerCP-Cy5-5	226	51.06	655	4	0.0081	42	Pass
Blue	A	PE-Cy7	130	100.54	711	6	0.0103	20	Pass
Red	C	APC	372	28.4	603	6	0.0203	160	Pass
Red	A	APC-Cy7	212	27.8	576	5	0.0046	567	Pass
Violet	B	V450	73	63.23	478	3	0.0403	1731	Pass
Violet	A	V500	112	34.21	488	4	0.0199	613	Pass

Specifications

PMTV Delta from baseline: 50 (BD)
Violet Laser Primary Channel Bright Bead %Robust CV: 6.00 (Recommended)
Blue Laser Primary Channel Bright Bead %Robust CV: 6.00 (Recommended)
Red Laser Primary Channel Bright Bead %Robust CV: 6.00 (Recommended)

Cytometer Performance Report

Cytometer: FACSCantoII
Cytometer Name: FACSCantoII
Serial Number: V33896201821
Input Device: Carousel
Tube Loaded Manually: Yes
Cytometer Configuration: Euroflow 3-laser, 8-color (4-2-2)

User: Administrator
Institution: N/A
Software: BD FACSDiva 8.0.2
Date: 10/03/2018 11:35 AM
Cytometer Baseline: 09/25/2018 11:17 AM
P/F: Pass

Laser Settings

Laser	Power (mW)	Power Spec. (mW)	Current (mA)	Current Guidelines (mA)	Delay (Trigger on FSC)	Delay (Trigger on Fluorescence)	Area Scaling Factor
Violet	59.55	47.98 - 71.98	N/A	N/A	-33.27	-32.69	1.08
Blue	20.37	16.27 - 24.41	1.27	0.99 - 1.49	0.00	0.00	1.50
Red	17.79	13.60 - 20.40	N/A	N/A	29.68	30.26	1.21

Window Extension: 7.00
FSC Area Scaling Factor: 0.92

Threshold

Threshold (FSC): 14926

Fluidics

FACSFlow Pressure: 3.90
Spec.: 3.80 - 4.00
Sample Pressure: 3.60

Comments